

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
2 °C	2.015 °C	2.3 °C	0.285 °C	± 0.5 °C	± 4.0E-01 °C
5 °C	4.983 °C	5.4 °C	0.417 °C	± 0.5 °C	± 4.0E-01 °C
8 °C	8.142 °C	8.5 °C	0.358 °C	± 0.5 °C	± 4.0E-01 °C