

<b>Nominal (Nominal)</b>	<b>Patrón (Standard)</b>	<b>UBP (UUT)</b>	<b>Error (Error)</b>	<b>Exactitud (Accuracy)</b>	<b>Incertidumbre (Uncertainty)</b>
22 °C	21.890 °C	22.3 °C	0.410 °C	± 0.5 °C	± 4.0E-01 °C
25 °C	24.868 °C	25.2 °C	0.332 °C	± 0.5 °C	± 4.0E-01 °C
31 °C	31.025 °C	30.8 °C	-0.225 °C	± 0.5 °C	± 4.0E-01 °C
48 %HR	47.431 %HR	51.5 %HR	4.069 %HR	± 5.0 %HR	± 1.9E+00 %HR
57 %HR	56.654 %HR	55.3 %HR	-1.354 %HR	± 5.0 %HR	± 1.9E+00 %HR
67 %HR	66.613 %HR	61.9 %HR	-4.713 %HR	± 5.0 %HR	± 1.9E+00 %HR