

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
22 °C	21.890 °C	22.3 °C	0.410 °C	± 0.5 °C	± 4.0E-01 °C
25 °C	24.868 °C	25.3 °C	0.432 °C	± 0.5 °C	± 4.0E-01 °C
31 °C	31.025 °C	31.1 °C	0.075 °C	± 0.5 °C	± 4.0E-01 °C
48 %HR	47.431 %HR	51.2 %HR	3.769 %HR	± 5.0 %HR	± 1.9E+00 %HR
57 %HR	53.586 %HR	56.1 %HR	2.514 %HR	± 5.0 %HR	± 1.9E+00 %HR
67 %HR	66.613 %HR	62.1 %HR	-4.513 %HR	± 5.0 %HR	± 1.9E+00 %HR