

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
22 °C	21.890 °C	22.1 °C	0.210 °C	± 0.5 °C	± 4.0E-01 °C
25 °C	24.868 °C	25.1 °C	0.232 °C	± 0.5 °C	± 4.0E-01 °C
31 °C	31.025 °C	30.8 °C	-0.225 °C	± 0.5 °C	± 4.0E-01 °C
48 %HR	47.431 %HR	51.8 %HR	4.369 %HR	± 5.0 %HR	± 1.9E+00 %HR
57 %HR	53.586 %HR	56.7 %HR	3.114 %HR	± 5.0 %HR	± 1.9E+00 %HR
67 %HR	66.613 %HR	63.9 %HR	-2.713 %HR	± 5.0 %HR	± 1.9E+00 %HR