

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
22 °C	21.890 °C	22.3 °C	0.410 °C	± 0.5 °C	± 4.0E-01 °C
25 °C	24.868 °C	25.3 °C	0.432 °C	± 0.5 °C	± 4.0E-01 °C
31 °C	31.025 °C	30.7 °C	-0.325 °C	± 0.5 °C	± 4.0E-01 °C
48 %HR	47.431 %HR	52.1 %HR	4.669 %HR	± 5.0 %HR	± 1.9E+00 %HR
57 %HR	56.654 %HR	57.0 %HR	0.346 %HR	± 5.0 %HR	± 1.9E+00 %HR
67 %HR	66.613 %HR	63.6 %HR	-3.013 %HR	± 5.0 %HR	± 1.9E+00 %HR