

| <b>Nominal (Nominal)</b> | <b>Patrón (Standard)</b> | <b>UBP (UUT)</b> | <b>Error (Error)</b> | <b>Exactitud (Accuracy)</b> | <b>Incertidumbre (Uncertainty)</b> |
|--------------------------|--------------------------|------------------|----------------------|-----------------------------|------------------------------------|
| 22 °C                    | 21.890 °C                | 22.2 °C          | 0.310 °C             | ± 0.5 °C                    | ± 4.0E-01 °C                       |
| 25 °C                    | 24.868 °C                | 25.2 °C          | 0.332 °C             | ± 0.5 °C                    | ± 4.0E-01 °C                       |
| 31 °C                    | 31.025 °C                | 30.8 °C          | -0.225 °C            | ± 0.5 °C                    | ± 4.0E-01 °C                       |
| 48 %HR                   | 47.431 %HR               | 52.4 %HR         | 4.969 %HR            | ± 5.0 %HR                   | ± 1.9E+00 %HR                      |
| 57 %HR                   | 56.654 %HR               | 61.4 %HR         | 4.746 %HR            | ± 5.0 %HR                   | ± 1.9E+00 %HR                      |
| 67 %HR                   | 66.613 %HR               | 64.6 %HR         | -2.013 %HR           | ± 5.0 %HR                   | ± 1.9E+00 %HR                      |