

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
22 °C	21.890 °C	21.9 °C	0.010 °C	± 0.5 °C	± 4.0E-01 °C
25 °C	24.868 °C	24.9 °C	0.032 °C	± 0.5 °C	± 4.0E-01 °C
31 °C	31.025 °C	30.7 °C	-0.325 °C	± 0.5 °C	± 4.0E-01 °C
48 %HR	47.431 %HR	52.0 %HR	4.569 %HR	± 5.0 %HR	± 1.9E+00 %HR
57 %HR	56.654 %HR	57.2 %HR	0.546 %HR	± 5.0 %HR	± 1.9E+00 %HR
67 %HR	66.613 %HR	62.8 %HR	-3.813 %HR	± 5.0 %HR	± 1.9E+00 %HR