

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
22 °C	21.980 °C	22.3 °C	0.320 °C	± 0.5 °C	± 4.0E-01 °C
25 °C	24.928 °C	25.2 °C	0.272 °C	± 0.5 °C	± 4.0E-01 °C
31 °C	30.453 °C	30.7 °C	0.247 °C	± 0.5 °C	± 4.0E-01 °C
48 %HR	46.961 %HR	45.3 %HR	-1.661 %HR	± 5.0 %HR	± 1.9E+00 %HR
57 %HR	56.511 %HR	53.1 %HR	-3.411 %HR	± 5.0 %HR	± 1.9E+00 %HR
67 %HR	66.613 %HR	61.9 %HR	-4.713 %HR	± 5.0 %HR	± 1.9E+00 %HR