

<b>Nominal (Nominal)</b>	<b>Patrón (Standard)</b>	<b>UBP (UUT)</b>	<b>Error (Error)</b>	<b>Exactitud (Accuracy)</b>	<b>Incertidumbre (Uncertainty)</b>
# -30 °C	-30.364 °C	-29.7 °C	0.664 °C	± 0.7 °C	± 4.0E-01 °C
-25 °C	-25.450 °C	-24.8 °C	0.650 °C	± 0.7 °C	± 4.0E-01 °C
-15 °C	-15.353 °C	-14.9 °C	0.453 °C	± 0.5 °C	± 4.0E-01 °C
-10 °C	-10.309 °C	-9.9 °C	0.409 °C	± 0.5 °C	± 4.0E-01 °C
2 °C	1.825 °C	2.3 °C	0.475 °C	± 0.5 °C	± 4.0E-01 °C
8 °C	7.741 °C	8.2 °C	0.459 °C	± 0.5 °C	± 4.0E-01 °C