

| <b>Nominal (Nominal)</b> | <b>Patrón (Standard)</b> | <b>UBP (UUT)</b> | <b>Error (Error)</b> | <b>Exactitud (Accuracy)</b> | <b>Incertidumbre (Uncertainty)</b> |
|--------------------------|--------------------------|------------------|----------------------|-----------------------------|------------------------------------|
| # -30 °C                 | -30.364 °C               | -30.1 °C         | 0.264 °C             | ± 0.7 °C                    | ± 4.0E-01 °C                       |
| -25 °C                   | -25.450 °C               | -25.2 °C         | 0.250 °C             | ± 0.7 °C                    | ± 4.0E-01 °C                       |
| -15 °C                   | -15.353 °C               | -15.1 °C         | 0.253 °C             | ± 0.5 °C                    | ± 4.0E-01 °C                       |
| -10 °C                   | -10.309 °C               | -10.2 °C         | 0.109 °C             | ± 0.5 °C                    | ± 4.0E-01 °C                       |
| 2 °C                     | 1.825 °C                 | 1.8 °C           | -0.025 °C            | ± 0.5 °C                    | ± 4.0E-01 °C                       |
| 8 °C                     | 7.741 °C                 | 7.8 °C           | 0.059 °C             | ± 0.5 °C                    | ± 4.0E-01 °C                       |