

<b>Nominal (Nominal)</b>	<b>Patrón (Standard)</b>	<b>UBP (UUT)</b>	<b>Error (Error)</b>	<b>Exactitud (Accuracy)</b>	<b>Incertidumbre (Uncertainty)</b>
# -30 °C	-30.364 °C	-30.1 °C	0.264 °C	± 0.7 °C	± 4.0E-01 °C
-25 °C	-25.450 °C	-25.1 °C	0.350 °C	± 0.7 °C	± 4.0E-01 °C
-15 °C	-15.353 °C	-15.1 °C	0.253 °C	± 0.5 °C	± 4.0E-01 °C
-10 °C	-10.309 °C	-10.3 °C	0.009 °C	± 0.5 °C	± 4.0E-01 °C
2 °C	1.825 °C	1.8 °C	-0.025 °C	± 0.5 °C	± 4.0E-01 °C
8 °C	7.741 °C	7.5 °C	-0.241 °C	± 0.5 °C	± 4.0E-01 °C