

| Patrón (Standard) | UBP (UUT) | Error (Error) | Exactitud (Accuracy) | Incertidumbre (Uncertainty) | |
|--------------------------|------------------|----------------------|-----------------------------|------------------------------------|--------------|
| 49.80 mmHg | 50 mmHg | 0.20 mmHg | ± 5 mmHg | ± | 6.6E-01 mmHg |
| 100.20 mmHg | 100 mmHg | -0.20 mmHg | ± 5 mmHg | ± | 6.5E-01 mmHg |
| 150.20 mmHg | 150 mmHg | -0.20 mmHg | ± 5 mmHg | ± | 6.6E-01 mmHg |
| 200.30 mmHg | 200 mmHg | -0.30 mmHg | ± 5 mmHg | ± | 6.8E-01 mmHg |
| 250.20 mmHg | 250 mmHg | -0.20 mmHg | ± 5 mmHg | ± | 7.5E-01 mmHg |
| 299.90 mmHg | 300 mmHg | 0.10 mmHg | ± 5 mmHg | ± | 7.5E-01 mmHg |