

Patrón (Standard)	UBP (UIT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)	
30.318 s	30.31 s	-0.008 s	-0.026%	± 0.01 %	± 9.2E-02 s
60.205 s	60.2 s	-0.005 s	-0.008%	± 0.01 %	± 9.2E-02 s
90.227 s	90.23 s	0.003 s	0.003%	± 0.01 %	± 9.2E-02 s
180.318 s	180.31 s	-0.008 s	-0.004%	± 0.01 %	± 9.2E-02 s
300.218 s	300.21 s	-0.008 s	-0.003%	± 0.01 %	± 9.2E-02 s
1200.105 s	1200.099 s	-0.006 s	0.000%	± 0.01 %	± 9.2E-02 s
3600.311 s	3600.305 s	-0.006 s	0.000%	± 0.01 %	± 9.2E-02 s