

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
30 °C	30.213 °C	29.9 °C	-0.313 °C	± 1.0 °C	± 4.0E-01 °C
30 %HR	30.345 %HR	34 %HR	3.655 %HR	± 5.0 %HR	± 1.9E+00 %HR
50 %HR	50.090 %HR	53 %HR	2.910 %HR	± 5.0 %HR	± 1.9E+00 %HR
70 %HR	69.578 %HR	74 %HR	4.422 %HR	± 5.0 %HR	± 1.9E+00 %HR