

<b>Nominal (Nominal)</b>	<b>Patrón (Standard)</b>	<b>UBP (UUT)</b>	<b>Error (Error)</b>	<b>Exactitud (Accuracy)</b>	<b>Incertidumbre (Uncertainty)</b>
30 °C	30.213 °C	30.3 °C	0.087 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	30.345 %HR	32.5 %HR	2.155 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	50.090 %HR	51.2 %HR	1.110 %HR	± 3.0 %HR	± 1.9E+00 %HR
70 %HR	69.578 %HR	69.7 %HR	0.122 %HR	± 3.0 %HR	± 1.9E+00 %HR