

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
30 °C	30.213 °C	30.2 °C	-0.013 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	30.345 %HR	32.9 %HR	2.555 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	50.090 %HR	52.2 %HR	2.110 %HR	± 3.0 %HR	± 1.9E+00 %HR
70 %HR	69.578 %HR	68.4 %HR	-1.178 %HR	± 3.0 %HR	± 1.9E+00 %HR