

Patrón (Standard)	UBP (UUT)	Error (Error)	E.M.P (Tolerance)	Incertidumbre (Uncertainty)	
23.20 μm	22.7 μm	-0.50 μm	$\pm 0.70 \mu\text{m}$	\pm	6.8E-02 μm
51.10 μm	50.4 μm	-0.70 μm	$\pm 1.5 \mu\text{m}$	\pm	9.7E-02 μm
124.60 μm	123.0 μm	-1.60 μm	$\pm 3.7 \mu\text{m}$	\pm	2.0E-01 μm
249.40 μm	247.0 μm	-2.40 μm	$\pm 7.5 \mu\text{m}$	\pm	3.9E-01 μm
478.00 μm	473.0 μm	-5.00 μm	$\pm 14.3 \mu\text{m}$	\pm	7.3E-01 μm
952.00 μm	945.0 μm	-7.00 μm	$\pm 28.6 \mu\text{m}$	\pm	1.5E+00 μm