

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
10 °C	9.917 °C	10.2 °C	0.283 °C	± 0.5 °C	± 4.0E-01 °C
15 °C	14.931 °C	14.7 °C	-0.231 °C	± 0.5 °C	± 4.0E-01 °C
25 °C	24.928 °C	24.8 °C	-0.128 °C	± 0.5 °C	± 4.0E-01 °C
32 °C	32.519 °C	32.6 °C	0.081 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	30.601 %HR	32.4 %HR	1.799 %HR	± 3.0 %HR	± 1.9E+00 %HR
70 %HR	69.987 %HR	67.6 %HR	-2.387 %HR	± 3.0 %HR	± 1.9E+00 %HR