

<b>Nominal (Nominal)</b>	<b>Patrón (Standard)</b>	<b>UBP (UUT)</b>	<b>Error (Error)</b>	<b>Exactitud (Accuracy)</b>	<b>Incertidumbre (Uncertainty)</b>
10 °C	9.917 °C	10.1 °C	0.183 °C	± 0.5 °C	± 4.0E-01 °C
15 °C	14.931 °C	14.8 °C	-0.131 °C	± 0.5 °C	± 4.0E-01 °C
25 °C	24.928 °C	24.8 °C	-0.128 °C	± 0.5 °C	± 4.0E-01 °C
32 °C	32.519 °C	32.4 °C	-0.119 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	30.601 %HR	32.7 %HR	2.099 %HR	± 3.0 %HR	± 1.9E+00 %HR
70 %HR	69.987 %HR	67.8 %HR	-2.187 %HR	± 3.0 %HR	± 1.9E+00 %HR