

<b>Patrón (Standard)</b>	<b>UBP (UUT)</b>	<b>Error (Error)</b>	<b>Exactitud (Accuracy)</b>	<b>Incertidumbre (Uncertainty)</b>
200.08 lx	203.2 lx	3.12 lx	± 20 lx	± 9.9E-02 lx
150.03 lx	153.1 lx	3.07 lx	± 15 lx	± 9.9E-02 lx
100.69 lx	103.7 lx	3.01 lx	± 10 lx	± 9.9E-02 lx
50.13 lx	53.3 lx	3.17 lx	± 5 lx	± 9.9E-02 lx