

<b>Patrón (Standard)</b>	<b>UBP (UUT)</b>	<b>Error (Error)</b>	<b>Exactitud (Accuracy)</b>	<b>Incertidumbre (Uncertainty)</b>	
200.1 lx	203 lx	-2.5 lx	± 20 lx	±	9.9E-02 lx
100.7 lx	103 lx	-2.4 lx	± 10 lx	±	6.7E-02 lx
50.1 lx	53 lx	-2.4 lx	± 5 lx	±	6.0E-02 lx