

Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)	
200.1 lx	203 lx	-3.3 lx	± 20 lx	±	9.9E-02 lx
100.7 lx	104 lx	-3.1 lx	± 10 lx	±	6.7E-02 lx
50.1 lx	54 lx	-3.6 lx	± 5 lx	±	6.0E-02 lx