

Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)	
200.08 lx	196.0 lx	-4.08 lx	± 20 lx	±	9.9E-02 lx
150.03 lx	146.1 lx	-3.93 lx	± 15 lx	±	9.9E-02 lx
100.69 lx	97.0 lx	-3.69 lx	± 10 lx	±	9.9E-02 lx
50.13 lx	46.3 lx	-3.83 lx	± 5 lx	±	9.9E-02 lx