

<b>Nominal (Nominal)</b>	<b>Patrón (Standard)</b>	<b>UBP (UUT)</b>	<b>Error (Error)</b>	<b>Exactitud (Accuracy)</b>	<b>Incertidumbre (Uncertainty)</b>
15 °C	14.931 °C	15.1 °C	0.169 °C	± 0.5 °C	± 4.0E-01 °C
25 °C	24.928 °C	25.3 °C	0.372 °C	± 0.5 °C	± 4.0E-01 °C
32 °C	32.519 °C	32.6 °C	0.081 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	29.558 %HR	32.3 %HR	2.742 %HR	± 3.0 %HR	± 1.9E+00 %HR
70 %HR	69.987 %HR	72.6 %HR	2.613 %HR	± 3.0 %HR	± 1.9E+00 %HR