

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
30 °C	30.453 °C	30.4 °C	-0.053 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	29.558 %HR	31.8 %HR	2.242 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	49.006 %HR	51.2 %HR	2.194 %HR	± 3.0 %HR	± 1.9E+00 %HR
75 %HR	75.100 %HR	75.6 %HR	0.500 %HR	± 3.0 %HR	± 1.9E+00 %HR