

Clase de Exactitud (Accuracy Class): Clase III (Class III)

---

Valor UBP	Valor Patrón	Error	EMP	Incertidumbre	Valor UBP	Valor Patrón	Error	EMP	Incertidumbre
m	m	mm	mm	(Uncertainty)	m	m	mm	mm	(Uncertainty)
0 a 0.8	0.79921	1.060	± 1.8	± 4.3E-01 mm					
0 a 0.9	0.89913	1.150	± 1.9	± 4.3E-01 mm					
0 a 1	0.99898	1.300	± 2.0	± 4.3E-01 mm					
0 a 1.1	1.09888	1.400	± 2.0	± 4.8E-01 mm					
0 a 1.2	1.19875	1.540	± 2.1	± 4.8E-01 mm					
0 a 1.3	1.29866	1.630	± 2.2	± 4.8E-01 mm					
0 a 1.4	1.39853	1.770	± 2.3	± 4.8E-01 mm					
0 a 1.5	1.49844	1.860	± 2.4	± 4.8E-01 mm					
0 a 1.6	1.59839	1.920	± 2.4	± 4.8E-01 mm					
0 a 1.7	1.69828	2.030	± 2.5	± 4.8E-01 mm					
0 a 1.8	1.79815	2.170	± 2.6	± 4.8E-01 mm					
0 a 1.9	1.89804	2.290	± 2.7	± 4.8E-01 mm					
0 a 2	1.99796	2.370	± 2.8	± 4.8E-01 mm					
0 a 2.1	2.09786	2.480	± 2.8	± 4.8E-01 mm					

---