

<b>Patrón (Standard)</b>		<b>UBP (UUT)</b>	<b>Error (Error)</b>	<b>Exactitud (Accuracy)</b>	<b>Incertidumbre (Uncertainty)</b>	
30 Hz	1.000 g-pk	1.003 g-pk	0.003 g-pk	± 0.01 g-pk	±	2.1E-02 g-pk
50 Hz	1.000 g-pk	1.001 g-pk	0.001 g-pk	± 0.01 g-pk	±	2.0E-02 g-pk
80 Hz	1.000 g-pk	1.003 g-pk	0.003 g-pk	± 0.01 g-pk	±	2.0E-02 g-pk
100 Hz	1.000 g-pk	0.999 g-pk	-0.001 g-pk	± 0.01 g-pk	±	2.1E-02 g-pk
500 Hz	1.000 g-pk	1.002 g-pk	0.002 g-pk	± 0.01 g-pk	±	2.0E-02 g-pk
1000 Hz	1.000 g-pk	1.001 g-pk	0.001 g-pk	± 0.01 g-pk	±	2.1E-02 g-pk
#2000 Hz	50.000 g-pk	50.003 g-pk	0.003 g-pk	± 0.01 g-pk	±	2.1E-02 g-pk
#5000 Hz	120.000 g-pk	120.003 g-pk	0.003 g-pk	± 0.01 g-pk	±	2.1E-02 g-pk