

<b>Nominal (Nominal)</b>	<b>Patrón (Standard)</b>	<b>UBP (UUT)</b>	<b>Error (Error)</b>	<b>Exactitud (Accuracy)</b>	<b>Incertidumbre (Uncertainty)</b>
30 °C	30.213 °C	30.5 °C	0.287 °C	± 1.0 °C	± 4.0E-01 °C
30 %HR	30.345 %HR	33 %HR	2.655 %HR	± 5.0 %HR	± 1.9E+00 %HR
50 %HR	50.090 %HR	53 %HR	2.910 %HR	± 5.0 %HR	± 1.9E+00 %HR
70 %HR	69.578 %HR	72 %HR	2.422 %HR	± 5.0 %HR	± 1.9E+00 %HR