

Clase de Exactitud (Accuracy Class): Clase II (Class II)

Valor UBP	Valor Patrón	Error	EMP	Incertidumbre	Valor UBP	Valor Patrón	Error	EMP	Incertidumbre
m	m	mm	mm	(Uncertainty)	m	m	mm	mm	(Uncertainty)
0 a 0.5	0.49933	0.670	± 0.8	± 4.3E-01 mm					
0 a 1	0.99912	0.880	± 1	± 4.3E-01 mm					
0 a 1.5	1.49931	0.690	± 1.2	± 4.8E-01 mm					
0 a 2	1.99921	0.790	± 1.4	± 4.8E-01 mm					
0 a 2.5	2.49920	0.800	± 1.6	± 4.8E-01 mm					
0 a 3	2.99914	0.860	± 1.8	± 4.8E-01 mm					
0 a 3.5	3.49947	0.530	± 2	± 4.8E-01 mm					
0 a 4	3.99942	0.580	± 2.2	± 4.8E-01 mm					
0 a 4.5	4.49936	0.640	± 2.4	± 4.8E-01 mm					
0 a 5	4.99935	0.650	± 2.6	± 4.8E-01 mm					
