

Clase de Exactitud (Accuracy Class): Clase II (Class II)

Valor UBP	Valor Patrón	Error	EMP	Incertidumbre (Uncertainty)	Valor UBP	Valor Patrón	Error	EMP	Incertidumbre (Uncertainty)
m	m	mm	mm		m	m	mm	mm	
0 a 0.1	0.100	0.33	± 0.6	± 4.3E-01 mm					
0 a 0.2	0.200	0.34	± 0.6	± 4.3E-01 mm					
0 a 0.3	0.300	0.35	± 0.7	± 4.3E-01 mm					
0 a 0.4	0.400	0.35	± 0.7	± 4.3E-01 mm					
0 a 0.5	0.500	0.36	± 0.8	± 4.3E-01 mm					
0 a 0.6	0.600	0.36	± 0.8	± 4.3E-01 mm					
0 a 0.7	0.700	0.37	± 0.8	± 4.3E-01 mm					
0 a 0.8	0.800	0.38	± 0.9	± 4.3E-01 mm					
0 a 0.9	0.900	0.39	± 0.9	± 4.3E-01 mm					
0 a 1	1.000	0.40	± 1	± 4.3E-01 mm					
0 a 1.2	1.200	0.43	± 1	± 4.7E-01 mm					
0 a 1.4	1.400	0.46	± 1.1	± 4.7E-01 mm					
0 a 1.6	1.600	0.49	± 1.2	± 4.7E-01 mm					
0 a 1.8	1.800	0.20	± 1.3	± 4.7E-01 mm					
0 a 2	2.000	0.22	± 1.4	± 4.7E-01 mm					
