

Patrón (Standard)	UBP (UUT)	Error (Error)	E.M.P (Tolerance)	Incertidumbre (Uncertainty)	
1.250 mm	1.250 mm	0 μm	$\pm 30 \mu\text{m}$	\pm	5.8E-01 μm
1.500 mm	1.500 mm	0 μm	$\pm 30 \mu\text{m}$	\pm	5.8E-01 μm
2.000 mm	2.000 mm	0 μm	$\pm 30 \mu\text{m}$	\pm	5.8E-01 μm
3.000 mm	3.000 mm	0 μm	$\pm 30 \mu\text{m}$	\pm	5.8E-01 μm
5.000 mm	5.000 mm	0 μm	$\pm 30 \mu\text{m}$	\pm	5.8E-01 μm
8.000 mm	8.000 mm	0 μm	$\pm 30 \mu\text{m}$	\pm	5.8E-01 μm
10.000 mm	10.000 mm	0 μm	$\pm 30 \mu\text{m}$	\pm	5.8E-01 μm
15.000 mm	15.000 mm	0 μm	$\pm 30 \mu\text{m}$	\pm	5.9E-01 μm
20.000 mm	20.001 mm	1 μm	$\pm 30 \mu\text{m}$	\pm	5.8E-01 μm
25.000 mm	25.002 mm	2 μm	$\pm 30 \mu\text{m}$	\pm	5.9E-01 μm