

Patrón (Standard)	UBP (UUT)	Error (Error)	E.M.P (Tolerance)	Incertidumbre (Uncertainty)	
23.2 μm	23.7 μm	0.5 μm	± 2.5 μm	±	5.8E-02 μm
51.1 μm	52.0 μm	0.9 μm	± 2.5 μm	±	5.8E-02 μm
125.1 μm	123 μm	-2.1 μm	± 3.8 μm	±	6.1E-02 μm
478.0 μm	481 μm	3.0 μm	± 14.3 μm	±	9.3E-02 μm
952 μm	955 μm	3.0 μm	± 28.6 μm	±	1.6E-01 μm