

Patrón (Standard)	UBP (UUT)	Error (Error)	E.M.P (Tolerance)	Incertidumbre (Uncertainty)
1 mm	0.97 mm	-30 μm	$\pm 10 \mu\text{m}$	$\pm 5.8\text{E}+00 \mu\text{m}$
1 mm	1.22 mm	-30 μm	$\pm 10 \mu\text{m}$	$\pm 5.8\text{E}+00 \mu\text{m}$
2 mm	1.47 mm	-30 μm	$\pm 10 \mu\text{m}$	$\pm 5.8\text{E}+00 \mu\text{m}$
2 mm	1.95 mm	-50 μm	$\pm 10 \mu\text{m}$	$\pm 5.8\text{E}+00 \mu\text{m}$
3 mm	2.94 mm	-60 μm	$\pm 10 \mu\text{m}$	$\pm 5.8\text{E}+00 \mu\text{m}$
5 mm	4.90 mm	-100 μm	$\pm 10 \mu\text{m}$	$\pm 5.8\text{E}+00 \mu\text{m}$
10 mm	9.80 mm	-200 μm	$\pm 10 \mu\text{m}$	$\pm 5.8\text{E}+00 \mu\text{m}$
15 mm	14.73 mm	-270 μm	$\pm 10 \mu\text{m}$	$\pm 5.8\text{E}+00 \mu\text{m}$
20 mm	19.60 mm	-400 μm	$\pm 10 \mu\text{m}$	$\pm 5.8\text{E}+00 \mu\text{m}$