

Patrón (Standard)		UBP (UIT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)	
7 Hz	0.400 g-pk	0.31 g-pk	-0.090 g-pk	± 0.06 g-pk	±	1.1E-02 g-pk
10 Hz	0.800 g-pk	0.90 g-pk	0.100 g-pk	± 0.12 g-pk	±	1.7E-02 g-pk
100 Hz	1.000 g-pk	0.90 g-pk	-0.100 g-pk	± 0.15 g-pk	±	2.1E-02 g-pk
2000 Hz	1.000 g-pk	0.90 g-pk	-0.100 g-pk	± 0.15 g-pk	±	2.2E-02 g-pk
5000 Hz	1.000 g-pk	0.90 g-pk	-0.100 g-pk	± 0.15 g-pk	±	3.2E-02 g-pk
10000 Hz	1.000 g-pk	0.90 g-pk	-0.100 g-pk	± 0.15 g-pk	±	4.1E-02 g-pk