

Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)	
49.70 mmHg	50 mmHg	0.30 mmHg	± 3 mmHg	±	6.6E-01 mmHg
99.80 mmHg	100 mmHg	0.20 mmHg	± 3 mmHg	±	6.5E-01 mmHg
149.90 mmHg	150 mmHg	0.10 mmHg	± 3 mmHg	±	6.6E-01 mmHg
199.80 mmHg	200 mmHg	0.20 mmHg	± 3 mmHg	±	6.8E-01 mmHg
249.80 mmHg	250 mmHg	0.20 mmHg	± 3 mmHg	±	7.5E-01 mmHg
299.70 mmHg	300 mmHg	0.30 mmHg	± 3 mmHg	±	7.5E-01 mmHg