

Patrón (Standard)	UBP (UUT)	Error (Error)	E.M.P (Tolerance)	Incertidumbre (Uncertainty)
1.000 mm	1.001 mm	1.000 μm	$\pm 20 \mu\text{m}$	$\pm 5.8\text{E-}01 \mu\text{m}$
1.250 mm	1.252 mm	2.000 μm	$\pm 20 \mu\text{m}$	$\pm 5.8\text{E-}01 \mu\text{m}$
1.500 mm	1.502 mm	2.000 μm	$\pm 20 \mu\text{m}$	$\pm 5.8\text{E-}01 \mu\text{m}$
2.000 mm	2.004 mm	4.000 μm	$\pm 20 \mu\text{m}$	$\pm 5.8\text{E-}01 \mu\text{m}$
3.000 mm	3.003 mm	3.000 μm	$\pm 20 \mu\text{m}$	$\pm 5.8\text{E-}01 \mu\text{m}$
5.000 mm	5.001 mm	1.000 μm	$\pm 20 \mu\text{m}$	$\pm 5.8\text{E-}01 \mu\text{m}$
8.000 mm	8.001 mm	1.000 μm	$\pm 20 \mu\text{m}$	$\pm 5.8\text{E-}01 \mu\text{m}$
10.000 mm	10.009 mm	9.000 μm	$\pm 20 \mu\text{m}$	$\pm 5.8\text{E-}01 \mu\text{m}$