

Patrón (Standard)	UBP (UUT)	Error (Error)	E.M.P (Tolerance)	Incertidumbre (Uncertainty)	
1.000 mm	1.001 mm	1.000 μm	$\pm 20 \mu\text{m}$	\pm	5.8E-01 μm
1.250 mm	1.258 mm	8.000 μm	$\pm 20 \mu\text{m}$	\pm	5.8E-01 μm
1.500 mm	1.506 mm	6.000 μm	$\pm 20 \mu\text{m}$	\pm	5.8E-01 μm
2.000 mm	2.004 mm	4.000 μm	$\pm 20 \mu\text{m}$	\pm	5.8E-01 μm
3.000 mm	3.003 mm	3.000 μm	$\pm 20 \mu\text{m}$	\pm	5.8E-01 μm
5.000 mm	5.002 mm	2.000 μm	$\pm 20 \mu\text{m}$	\pm	5.8E-01 μm
8.000 mm	8.003 mm	3.000 μm	$\pm 20 \mu\text{m}$	\pm	5.8E-01 μm
10.000 mm	10.006 mm	6.000 μm	$\pm 20 \mu\text{m}$	\pm	5.8E-01 μm