

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
23 °C	23.434 °C	23.4 °C	-0.034 °C	± 1.0 °C	± 4.0E-01 °C
30 %HR	29.558 %HR	29.5 %HR	-0.058 %HR	± 5.0 %HR	± 1.9E+00 %HR
50 %HR	49.006 %HR	46.5 %HR	-2.506 %HR	± 5.0 %HR	± 1.9E+00 %HR
75 %HR	75.100 %HR	73.5 %HR	-1.600 %HR	± 5.0 %HR	± 1.9E+00 %HR