

<b>Nominal</b> ( <i>Nominal</i> )	<b>Patrón</b> ( <i>Standard</i> )	<b>UBP</b> ( <i>UUT</i> )	<b>Error</b> ( <i>Error</i> )	<b>Exactitud</b> ( <i>Accuracy</i> )	<b>Incertidumbre</b> ( <i>Uncertainty</i> )
30 °C	30.173 °C	30.3 °C	0.127 °C	± 1.0 °C	± 4.0E-01 °C
30 %HR	28.055 %HR	29 %HR	0.945 %HR	± 5.0 %HR	± 1.9E+00 %HR
50 %HR	51.122 %HR	51 %HR	-0.122 %HR	± 5.0 %HR	± 1.9E+00 %HR
70 %HR	64.415 %HR	65 %HR	0.585 %HR	± 5.0 %HR	± 1.9E+00 %HR