

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
30 °C	30.173 °C	30.6 °C	0.427 °C	± 1.0 °C	± 4.0E-01 °C
30 %HR	28.055 %HR	31 %HR	2.945 %HR	± 5.0 %HR	± 1.9E+00 %HR
50 %HR	51.122 %HR	54 %HR	2.878 %HR	± 5.0 %HR	± 1.9E+00 %HR
70 %HR	64.415 %HR	62 %HR	-2.415 %HR	± 5.0 %HR	± 1.9E+00 %HR