

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
-25 °C	-24.889 °C	-24.8 °C	0.089 °C	± 0.7 °C	± 4.0E-01 °C
0 °C	0.080 °C	0.5 °C	0.420 °C	± 0.5 °C	± 4.0E-01 °C
5 °C	5.294 °C	5.7 °C	0.406 °C	± 0.5 °C	± 4.0E-01 °C
10 °C	10.308 °C	10.7 °C	0.392 °C	± 0.5 °C	± 4.0E-01 °C
25 °C	25.450 °C	25.9 °C	0.450 °C	± 0.5 °C	± 4.0E-01 °C
30 °C	30.463 °C	30.9 °C	0.437 °C	± 0.5 °C	± 4.0E-01 °C
40 °C	40.090 °C	40.5 °C	0.410 °C	± 0.5 °C	± 4.0E-01 °C

