

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
-25 °C	-24.889 °C	-24.9 °C	-0.011 °C	± 0.7 °C	± 4.0E-01 °C
0 °C	0.080 °C	0.4 °C	0.320 °C	± 0.5 °C	± 4.0E-01 °C
5 °C	5.294 °C	5.6 °C	0.306 °C	± 0.5 °C	± 4.0E-01 °C
10 °C	10.308 °C	10.6 °C	0.292 °C	± 0.5 °C	± 4.0E-01 °C
25 °C	25.450 °C	25.8 °C	0.350 °C	± 0.5 °C	± 4.0E-01 °C
30 °C	30.463 °C	30.8 °C	0.337 °C	± 0.5 °C	± 4.0E-01 °C
40 °C	40.090 °C	40.4 °C	0.310 °C	± 0.5 °C	± 4.0E-01 °C

