

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
-25 °C	-24.789 °C	-24.6 °C	0.189 °C	± 0.7 °C	± 4.0E-01 °C
0 °C	0.381 °C	0.7 °C	0.319 °C	± 0.5 °C	± 4.0E-01 °C
5 °C	5.094 °C	5.5 °C	0.406 °C	± 0.5 °C	± 4.0E-01 °C
10 °C	10.007 °C	10.4 °C	0.393 °C	± 0.5 °C	± 4.0E-01 °C
25 °C	25.149 °C	25.5 °C	0.351 °C	± 0.5 °C	± 4.0E-01 °C
30 °C	30.263 °C	30.5 °C	0.237 °C	± 0.5 °C	± 4.0E-01 °C
40 °C	39.990 °C	40.3 °C	0.310 °C	± 0.5 °C	± 4.0E-01 °C

