

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
-25 °C	-24.989 °C	-24.8 °C	0.189 °C	± 0.7 °C	± 4.0E-01 °C
0 °C	0.280 °C	0.6 °C	0.320 °C	± 0.5 °C	± 4.0E-01 °C
5 °C	4.993 °C	5.4 °C	0.407 °C	± 0.5 °C	± 4.0E-01 °C
10 °C	9.907 °C	10.3 °C	0.393 °C	± 0.5 °C	± 4.0E-01 °C
25 °C	25.049 °C	25.4 °C	0.351 °C	± 0.5 °C	± 4.0E-01 °C
30 °C	30.163 °C	30.5 °C	0.337 °C	± 0.5 °C	± 4.0E-01 °C
40 °C	39.990 °C	40.2 °C	0.210 °C	± 0.5 °C	± 4.0E-01 °C

