

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
30 %HR	30.877 %HR	28.6 %HR	-2.277 %HR	± 3 %HR	± 1.9E+00 %HR
50 %HR	50.509 %HR	48.8 %HR	-1.709 %HR	± 3 %HR	± 1.9E+00 %HR
70 %HR	70.754 %HR	70.0 %HR	-0.754 %HR	± 3 %HR	± 1.9E+00 %HR
90 %HR	90.540 %HR	92.2 %HR	1.660 %HR	± 5 %HR	± 1.9E+00 %HR

