

| Nominal (Nominal) | Patrón (Standard) | UBP (UUT) | Error (Error) | Exactitud (Accuracy) | Incertidumbre (Uncertainty) |
|-------------------|-------------------|-----------|---------------|----------------------|-----------------------------|
| -25 °C | -24.989 °C | -24.6 °C | 0.389 °C | ± 1.0 °C | ± 4.0E-01 °C |
| 0 °C | 0.280 °C | 0.6 °C | 0.320 °C | ± 0.45 °C | ± 4.0E-01 °C |
| 5 °C | 4.993 °C | 5.3 °C | 0.307 °C | ± 0.45 °C | ± 4.0E-01 °C |
| 10 °C | 9.907 °C | 10.2 °C | 0.293 °C | ± 0.45 °C | ± 4.0E-01 °C |
| 25 °C | 25.049 °C | 25.3 °C | 0.251 °C | ± 0.45 °C | ± 4.0E-01 °C |
| 30 °C | 30.163 °C | 30.4 °C | 0.237 °C | ± 0.45 °C | ± 4.0E-01 °C |
| 40 °C | 40.090 °C | 40.3 °C | 0.210 °C | ± 0.45 °C | ± 4.0E-01 °C |

