

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
-25 °C	-24.989 °C	-24.6 °C	0.389 °C	± 0.7 °C	± 4.0E-01 °C
0 °C	0.381 °C	0.8 °C	0.419 °C	± 0.5 °C	± 4.0E-01 °C
5 °C	5.094 °C	5.4 °C	0.306 °C	± 0.5 °C	± 4.0E-01 °C
10 °C	10.107 °C	10.4 °C	0.293 °C	± 0.5 °C	± 4.0E-01 °C
25 °C	25.049 °C	25.3 °C	0.251 °C	± 0.5 °C	± 4.0E-01 °C
30 °C	30.062 °C	30.3 °C	0.238 °C	± 0.5 °C	± 4.0E-01 °C
40 °C	39.990 °C	40.2 °C	0.210 °C	± 0.5 °C	± 4.0E-01 °C

