

<b>Nominal</b> ( <i>Nominal</i> )	<b>Patrón</b> ( <i>Standard</i> )	<b>UBP</b> ( <i>UUT</i> )	<b>Error</b> ( <i>Error</i> )	<b>Exactitud</b> ( <i>Accuracy</i> )	<b>Incertidumbre</b> ( <i>Uncertainty</i> )
10 °C	9.606 °C	9.9 °C	0.294 °C	± 0.5 °C	± 4.0E-01 °C
25 °C	25.600 °C	25.9 °C	0.300 °C	± 0.5 °C	± 4.0E-01 °C
40 °C	40.301 °C	40.6 °C	0.299 °C	± 0.5 °C	± 4.0E-01 °C
20 %HR	18.648 %HR	20.6 %HR	1.952 %HR	± 3.0 %HR	± 1.9E+00 %HR
60 %HR	57.953 %HR	60.3 %HR	2.347 %HR	± 3.0 %HR	± 1.9E+00 %HR
75 %HR	75.059 %HR	77.6 %HR	2.541 %HR	± 3.0 %HR	± 1.9E+00 %HR