

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
10 °C	9.606 °C	10.0 °C	0.394 °C	± 0.5 °C	± 4.0E-01 °C
25 °C	25.600 °C	25.9 °C	0.300 °C	± 0.5 °C	± 4.0E-01 °C
40 °C	40.301 °C	40.2 °C	-0.101 °C	± 0.5 °C	± 4.0E-01 °C
20 %HR	18.648 %HR	20.7 %HR	2.052 %HR	± 3.0 %HR	± 1.9E+00 %HR
60 %HR	57.953 %HR	59.8 %HR	1.847 %HR	± 3.0 %HR	± 1.9E+00 %HR
75 %HR	75.059 %HR	76.3 %HR	1.241 %HR	± 3.0 %HR	± 1.9E+00 %HR