

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
10 °C	9.606 °C	9.8 °C	0.194 °C	± 0.5 °C	± 4.0E-01 °C
25 °C	25.600 °C	25.5 °C	-0.100 °C	± 0.5 °C	± 4.0E-01 °C
40 °C	40.301 °C	40.1 °C	-0.201 °C	± 0.5 °C	± 4.0E-01 °C
20 %HR	18.648 %HR	20.8 %HR	2.152 %HR	± 3.0 %HR	± 1.9E+00 %HR
60 %HR	57.953 %HR	60.0 %HR	2.047 %HR	± 3.0 %HR	± 1.9E+00 %HR
75 %HR	75.059 %HR	76.7 %HR	1.641 %HR	± 3.0 %HR	± 1.9E+00 %HR